

Fresh Concrete Testing

54

V-Funnel Apparatus



S/S 10 Li V-Funnel 54-C0147 /S/S.CON

GENERAL DESCRIPTION

Used to determine the V-funnel flow time. The test is not suitable when the maximum size of the aggregate exceeds 20 mm.

The apparatus consists of a stainless steel V shaped frame fit with a watertight magnetically held gate and supported by a plated mild steel frame assure the top funnel horizontal.

ORDERING INFO

54-C0147/S/S.CON V-Funnel apparatus



53 Granite Street, GEEBUNG, QLD, 4034, Australia | P.O. Box 124, VIRGINIA, QLD 4014, Australia

Ph.: +61 (0)7 3265 4952 | Fax: +61 (0)7 3265 2713 | Email: sales@geo-con.com.au | Web: www.geo-con.com.au